



REPUBLIC OF LEBANON  
MINISTRY OF  
TELECOMMUNICATIONS

General directorate of exploitation & maintenance  
General director

To: DEKRA Testing and certification, S.A.U.

Our reference: 11266 /E&M/2017

Date: 14/11/2017

Your reference: 54411

Date: 03 Aug. 2017

Subject: Approval of Accessory to hearing instrument.

Responding your application submitted on behalf of "HIMSA II K/S," requesting our approval for the following product:

Product model: CPD-1

Applicant: HIMSA II K/S,

Lyngbyvej 28, 1. Sal, DK- 2100 Copenhagen Ø, Denmark.

Manufacturer: Same as applicant.

Consultant: DEKRA Testing and certification, S.A.U.

Parque Tecnológico de Andalucía, c/Severo Ochoa N° 2 – 29590 –  
Campanillas, Málaga, Spain

Frequencies: 2402 – 2480 MHz (Bluetooth LE); 2404 – 2478 MHz (Proximity)

RF Power: <5 dBm EIRP (Bluetooth LE); <17 dBm EIRP (Proximity)

Description: Wireless accessory to hearing instrument supporting Bluetooth LE and proprietary 2.4 GHz protocol called Proximity.

We would like to inform you that, unless otherwise indicated, the mentioned above product is technically approved to be used in our country.

Sincerely yours.

General Director of Exploitation  
And Maintenance

Engr. Bassel Al Ayoubi

