



DEKRA Testing and Certification, S.A.U.

Product certification body accredited by ENAC with accreditation No. 134/C-PR301
and designated by the competent national authority of Spain

to act as Notified Body (Notified Body No: 1909) in accordance with the Directive 2014/53/EU of 16 April 2014

Directive 2014/53/EU – EU-TYPE EXAMINATION CERTIFICATE

Identification Number: **76170RNB.001**
Issue date: **2023-12-19**

MANUFACTURER DETAILS:

Company name: **HIMSA II K/S**
Address: **Lyngbyvej 28, 1. DK-2100 Copenhagen Ø
Denmark**

EQUIPMENT DETAILS:

Type of equipment: **Hearing instrument programming interface**
Brand name: **Noahlink Wireless 2**
Model name: **CPD-2**
HW version: **NLW2-PS**
SW version: **Test FW w HCI**

SCOPE OF OPINION:

Essential requirements	Specifications / Standards	Submitted documents
Article 3.1(a): Electrical safety	EN 62368-1:2014 + AC:2015 + AC:2017 + A11:2017	Test report
Article 3.1(a): EMF exposure	IEC 62479:2010	Test report
Article 3.1(b): EMC	EN 301 489-1 V2.2.3 Draft EN 301 489-17 V3.2.6	Test report
Article 3.2: Radio spectrum use	EN 300 328 v2.2.2	Test report

OPINION:

Our opinion in accordance with Annex III of DIRECTIVE 2014/53/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 16 April 2014 on radio equipment and the mutual recognition of their conformity is that the equipment complies with the requirements of that directive stated in the above scope.

This opinion has 1 annex with 1 page and it is only valid in conjunction with it.

Signed on behalf of DEKRA Testing and Certification, S.A.U. in Málaga (Spain)

Name: **Ricardo Orejas**
Position: **Responsible for Certification**



Annex I to EU-Type Examination Certificate No. 76170RNB.001

TECHNICAL DOCUMENTATION:

Held at: HIMSA II K/S
Address: Lyngbyvej 28, 1. DK-2100 Copenhagen Ø
Denmark

TECHNICAL FEATURES AND CHARACTERISTICS:

Operating frequency bands: BTLE: 2400 – 2483.5 MHz
Modulations: BTLE: GFSK
Output power: BTLE: 7.1 dBm EIRP
Antenna: Internal antenna. Gain = 3 dBi
Operating voltage: 5 VDC
Operating temperature: 0°C to +55°C
Intended use: Wireless data interface to hearing instrument

CONFORMITY DETAILS:

Essential requirements	Specifications / Standards	Reference documents
Article 3.1(a): Electrical safety	EN 62368-1:2014 + AC:2015 + AC:2017 + A11:2017	76170RSE.001
Article 3.1(a): EMF exposure	IEC 62479:2010	76170RAN.001A1
Article 3.1(b): EMC	EN 301 489-1 V2.2.3 Draft EN 301 489-17 V3.2.6	76170REM.001
Article 3.2: Radio spectrum use	EN 300 328 v2.2.2	76170RRF.001

REMARKS AND COMMENTS:

It is mandatory to inform DEKRA Testing and Certification, S.A.U. in writing about any change in the approved equipment identified in this certificate, which could affect the conformity of the apparatus with the essential requirements or the conditions of validity of this certificate.